

**Intertek**

# Certificate of Compliance

TO THE LOW VOLTAGE DIRECTIVE 2006/95/EC

*Project No.:* DL07000017-71CA  
*Type of equipment:* MICROSCOPE  
*Document holder:* **MAGNUS ANALYTICS**  
A-5 Mohan Co-operative Industrial Estate  
Mathura Road, New Delhi – 110 044  
India  
*Manufacturer:* **MAGNUS ANALYTICS**  
A-5 Mohan Co-operative Industrial Estate  
Mathura Road, New Delhi – 110 044  
India  
*Type designation:* MLX (Monocular/Binocular/Trinocular )  
MLXi (Binocular/Trinocular)  
*Technical data:* Voltage:220-240V (AC) 50 Hz rated Current Input 0.2Amp  
*Trademark:* **Magnus**

This Notice of Presumption of Conformity has been based on the following:

CE Based on Review of Safety test report issued by TUV with reference number P279/1026/Q/06062004

**Applicable standard:**

CE IEC 61010-1 :90 + A1 :92 + A2 :95

CE CISPER-22, IEC 61000-4-2,4,5,11 IEC 61000-3-2, 3

**Other Info:**

CE Manufacturer ensures that the manufactured products will be identical to the samples evaluated

We confirm that the samples evaluated, according to our review of the above test reports and with our complementary evaluation, without any reasonable doubt will fulfil the requirements concerning electrical safety according to above mentioned standards harmonised with the Low Voltage Directive 2006/95/EC & EMC Directive 89/336/EC. Obvious electrical safety risks other than those covered by the above standards were not revealed during our evaluations.

Note: This certificate supersedes the issue dated 5<sup>th</sup> June 2008 with the manufacturer name from Olympus (India) Pvt. Ltd. to Magnus analytics.

**The product was found to contain the required items for CE Markings according to:**

**The Low Voltage Directive 2006/95/EC**

**The EMC Directive 89/336/EC**

*Note: This Verification is for the exclusive use of Intertek's Client and is provided pursuant to the agreement between Intertek and its Client. Intertek's responsibility and liability are limited to the terms and conditions of the agreement. Intertek assumes no liability to any party, other than to the Client in accordance with the agreement, for any loss, expense or damage occasioned by the use of this Verification. Only the Client is authorized to copy or distribute this Verification. Any use of the Intertek name or one of its marks for the sale or advertisement of the tested material, product or service must first be approved in writing by Intertek. The observations and inspection results referenced from this Verification are relevant only to the sample inspected. This Verification by itself does not imply that the material, product, or service is or has ever been under an Intertek certification program.*

**Intertek India Private Limited**  
E 20 Block B 1 Mohan Cooperative Industrial Area  
Mathura Road, New Delhi – 110 044, India  
Tel: + 91 11 4159 5400 Fax: +91 11 4159 5475  
<http://www.intertek.com>

  
**Rakesh Chaurasia**  
April 18<sup>th</sup>, 2009

This certificate may only be reproduced in its entirety and without any change.